



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: 42275

Description: Datalogger, Temperature + Humidity

Date of Issue: 23-May-17

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: W6M21609-16242-E-11
Report Date of Issue: 3/24/2017

Standards:

EN 61326-1:2013
CISPR11
IEC/EN 61000-4-2 (2009)/-3 (2006+A2:2010)/-8 (2010)

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive and RoHS Directive.

A handwritten signature in black ink, appearing to read "Mark Sultzbach", written over a horizontal line.

Mark Sultzbach / QA Administrator
(for Tony Campagna / Director of Quality Services)