

### Extech RH550-P 探头和 RH550-C 线缆

#### Copyright

© 2020, FLIR Systems, Inc.

All rights reserved worldwide. Names and marks appearing herein are either registered trademarks or trademarks of FLIR Systems and/or its subsidiaries. All other trademarks, trade names or company names referenced herein are used for identification only and are the property of their respective owners.

#### Document identity

Publ. No.: NAS100036  
Release: AA  
Commit: 63972  
Head: 63972  
Language: zh-CN  
Modified: 2020-02-19  
Formatted: 2020-02-19

#### Website

<http://www.extech.com>  
<http://www.flir.com>

#### Customer support

<http://support.flir.com>

#### Disclaimer

Specifications subject to change without further notice. Models and accessories subject to regional market considerations. License procedures may apply. Products described herein may be subject to US Export Regulations. Please refer to [exportquestions@flir.com](mailto:exportquestions@flir.com) with any questions.

## 目录

1 RH550-P 替换探头 .....	1
2 RH550-C 延长线 .....	1
3 保修 .....	1
4 客户支持 .....	1

## 1 RH550-P 替换探头

RH550-P 温度和湿度探头是 Extech RH550 温度和湿度仪表的替换探头。由于 RH550-P 具有固定的校准标准，因此可以确保 RH550 仪表的准确性。

## 2 RH550-C 延长线

用户使用 RH550-C 探头延长线能够让探头远离 RH550 主机。同时还可以串联多达 3 条线缆，而不会有损温度和湿度读数的准确性。线缆长度约为 197 厘米 (6.5 英尺)。包括两端连接器在内的线缆长度约为 210 厘米 (6.9 英尺)。

## 3 保修

FLIR Systems, Inc. 保证此 Extech 品牌探头和线缆自发货之日起六个月内无零件和工艺缺陷。

## 4 客户支持

客户支持电话：美国 (866) 477-3687；国际 +1 (603) 324-7800

校准、维修和退回电子邮件地址：[repair@extech.com](mailto:repair@extech.com)

技术支持：<https://support.flir.com>

Extech 主网站：[www.extech.com](http://www.extech.com)