



**FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND**  
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

## ***Declaration of Conformity***

**Extech Model: MA150**

**Description: 200A AC Mini Clamp Meters+ Voltage Detector**

**Date of Issue: 30-Mar-15**

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

**EMC Directive: 2004/108/EC**  
**Report Number: WT088000955**  
**Report Date of Issue: 5/20/2008**

**Standards:**

EN 61326-1:2006  
EN 61326-2-1:2006  
IEC 61326-1:2005  
IEC 61326-2-1:2005

**LVD Directive 2006/95/EC**  
**Report Number: WT133004552**  
**Report Date of Issue: 4/29/2014**

**Standards:**

EN 61010-1:2010  
EN 61010-2-032:2012  
EN 61010-2-033:2012

**RoHS Directive: 2011/65/EU**

**Standard:**

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, and RoHS Directive.

---

Tony Campagna  
Director of Quality/Services